

# EXFORMMA TECHNOLOGIES (HK) LTD



Tel: 852-91930131  
Email: exformma@gmail.com

**Product Name: Disposable Mask (M005)**

(Filtration Efficiency  $\geq$  95%)

## Description:

<b>Type:</b>	Single-core nose wire; double-core nose wire; whole plastic nose wire
<b>Standard:</b>	YY/T 0969-2013
	Anti-Dust, Anti-Virus, PM2.5
<b>Application:</b>	Offices, hospitals, hotels and many more places
<b>Function:</b>	3 ply, 2 <sup>nd</sup> layer filter BFE 99%

## Test:

Residual ethylene oxide ( $\mu\text{g} / \text{g}$ )	Test Method	Standard value and tolerance	Result
Bacterial filtration efficiency (%)	GB/T 14233.1-2008	$\leq 10$	Not detected (detection limit: 0.5)
	YY 0469-2011 Appendix B Test strain: Staphylococcus aureus ATCC 6538 Test area: 40cm <sup>2</sup> Gas flow rate: 28.3L / min Average particle diameter: 3.0 $\mu\text{m}$ Positive QC value: 1.9 $\times$ 10 <sup>3</sup> CFU Negative QC value: <1CFU	$\geq 95$	BFE 99.7 BFE 99.6 BFE 99.7

## Product Photo:

